What Is The Predominant Source of Intravascular Catheter Infections?

Leonard A. Mermel^{1,2}

¹Division of Infectious Diseases, Rhode Island Hospital, and ²Department of Medicine, Warren Alpert Medical School of Brown University, Providence, Rhode Island

The predominant source of intravascular catheter-related bloodstream infections has been a research and clinical question for more than 30 years. During that time, we've moved from the position of a single source predominating in all clinical scenarios to a more realistic appraisal that both skin at the insertion site and the catheter hub/connector (ie, an extraluminal and an intraluminal source of infection, respectively) are important and that maximally effective prevention programs must address both sources of infection.

Debate has existed regarding the major route whereby microbes infect intravascular catheters, an extraluminal route arising from skin at the catheter insertion site or an intraluminal route emanating from the catheter hub, from the catheter tubing connection, or less frequently, from contaminated intravenous fluids. Most of the evidence suggests that, in general, an extraluminal source of infection predominates in catheters placed for a shorter duration of time, whereas an intraluminal source predominates with more prolonged dwell times. In a study of 25 short-term catheter-related bloodstream infections (CRBSIs) that used molecular fingerprinting to confirm the source of infection, 15 CRBSIs were from an extraluminal source, 3 CRBSIs were

Clinical Infectious Diseases 2011;52(2):211–212

© The Author 2011 Published by Oxford University Press on behalf of the Infectious Diseases Society of America. All rights reserved. For Permissions, please e-mail: journals. permissions@oup.com. 1058-4038/2011/522-0001\$37.00

DOI: 10.1093/cid/ciq108

from an intraluminal source, and 7 CRBSIs could have originated from either source [1]. In sharp contrast, a study of 24 long-term CRBSIs found 5 CRBSIs were from an extraluminal source, 16 CRBSIs were from an intraluminal source, 2 CRBSIs could have arisen from either source, and 1 case resulted from hematogenous seeding of the catheter [2]. In another study that used molecular fingerprinting, the median duration of catheterization was 14 d for CRBSIs with a confirmed extraluminal source, 24 d when the source was either extraluminal or intraluminal, and 64 d when the source was confirmed as intraluminal [3]. Intraluminal colonization with biofilm-producing microbes was more widespread after prolonged catheterization in a study that found >40% of the intraluminal surface of catheters in situ after >30 d was covered with biofilm compared with only 15% of the intraluminal surface of catheters in place for <10 d (P < .001) [4]. Catheter-drawn blood cultures were performed weekly in a study of patients with hemodialysis catheters [5]. If a catheter-drawn blood culture revealed microbial growth, then weekly,

percutaneously drawn blood cultures were obtained. Of 31 patients, 21 developed positive catheter-drawn blood cultures at a mean dwell time of 27 d. Of these 21 patients, 12 went on to develop concordant microbial growth from percutaneously drawn blood cultures at a mean time of 32 d after the first positive catheter-drawn blood cultures were obtained. These findings suggest that intraluminal catheter colonization, as measured by positive blood cultures drawn through the catheter, occurs in long-term catheters, and if left unchecked, this can lead to true CRBSI.

CRBSIs arising from the insertion site are extraluminal, and adequate cutaneous antisepsis of the insertion site reduces risk of such infections [6]. CRBSIs arising from an intraluminal source reflects a breach in aseptic technique when manipulating catheter hubs, caps, connectors, or stopcocks, or contamination of the infusate itself. In 1 study, 31% of nurses did not disinfect needleless catheter connectors before accessing them and 17% of "discarded" blood samples from blood drawn through these needleless connectors had microbial growth

Received 7 September 2010; accepted 8 October 2010; Correspondence: Leonard Mermel, DO, ScM, AM (Hon), FACP, FIDSA, FSHEA, Division of Infectious Diseases, Rhode Island Hospital, 593 Eddy St., Providence, RI 02903 (Imermel@lifespan.org).

[7]. This may reflect inadequate time to properly clean the connectors before accessing them, inadequate training, or both. Likewise, a low nurse-to-patient ratio independently increases the risk of catheter infection [8], likely reflecting lapses in aseptic technique. Similarly, insufficiently trained nurses working in an intensive care unit setting independently increases the risk of such infections [9].

Some needleless connectors are associated with an increased incidence of catheter infection, likely from an intraluminal source [7, 10]. This may be due to of difficulty cleaning the surface of some currently marketed connectors, the inability to disinfect the internal connector components, or behavioral issues, as noted above, leading to suboptimal disinfection of connectors prior to their being accessed, or no disinfection at all. On the other hand, a decreased risk of catheter colonization with use of needleless connectors has been found in some studies [11, 12], suggesting that there are likely differences in the risk of intraluminal contamination and resultant bloodstream infection among the devices now in clinical use.

In conclusion, both extraluminal and intraluminal routes of infection are important in the pathogenesis of central venous catheter-related infections. Soon after insertion, the extraluminal route of infection predominates, whereas the intraluminal route does so after a more extended dwell time. Thus, a focus on catheter insertion will help prevent CRBSIs that occur within days of catheterization and a focus on catheter maintenance will help prevent later infections. Behavioral changes reflecting education, evidence-based catheter insertion, and maintenance bundle use [13-15], as well as the promise of novel catheter and connector advances [16, 17], antimicrobial dressings [18], and catheter flush solutions [19] will further reduce risk of catheter infections.

Acknowledgments

Partial, unrestricted support was provided by Catheter Connections, which had no influence on the content of this article.

Finanical support. This work was supported by Catheter Connections.

Potential conflicts of interest. L.A.M. has received research support from Theravance and Pfizer, and he has served as a consultant for CorMedix, Ash Access, Semprus, CareFusion, Catheter Connections, and Sage.

References

- Safdar N, Maki DG. The pathogenesis of catheter-related bloodstream infection with noncuffed short-term central venous catheters. Intensive Care Med 2004; 30:62–7.
- Segura M, Lladó L, Guirao X, et al. A prospective study of a new protocol for 'in situ' diagnosis of central venous catheter related bacteraemia. Clin Nutr **1993**; 12:103–7.
- 3. Douard MC, Clementi E, Arlet G, et al. Negative catheter-tip culture and diagnosis of catheter-related bacteremia. Nutrition **1994**; 10:397–404.
- Raad I, Costerton W, Sabharwal U, Sacilowski M, Anaissie E, Bodey GP. Ultrastructural analysis of indwelling vascular catheters: A quantitative relationship between luminal colonization and duration of placement. J Infect Dis **1993**; 168:400–7.
- Dittmer ID, Sharp D, McNulty CA, Williams AJ, Banks RA. A prospective study of central venous hemodialysis catheter colonization and peripheral bacteremia. Clin Nephrol 1999; 51:34–9.
- Chaiyakunapruk N, Veenstra DL, Lipsky BA, Saint S. Chlorhexidine compared with povidone-iodine solution for vascular cathetersite care: A meta-analysis. Ann Intern Med 2002; 136:792–801.
- Karchmer TB, Cook EM, Palavecino E, et al. Needleless valve ports may be associated with a high rate of catheter-related bloodstream infection [abstract]. In: Program of the 15thAnnual Meeting of the Society for Healthcare Epidemiology of America (Los Angeles, CA). Washington, DC: Society for Healthcare Epidemiology of America, 2005. 307.
- 8. Fridkin SK, Pear SM, Williamson TH, Galgiani JN, Jarvis WR. The role of

understaffing in central venous catheterassociated bloodstream infections. Infect Control Hosp Epidemiol **1996**; 17:150–8.

- Alonso-Echanove J, Edwards JR, Richards MJ, et al. Effect of nurse staffing antimicrobialimpregnated central venous catheters on the risk for bloodstream infections in intensive care units. Infect Control Hosp Epidemiol 2003; 24:916–25.
- Jarvis WR, Murphy C, Hall KK, et al. Health care-associated bloodstream infections associated with negative- or positive-pressure or displacement mechanical valve needleless connectors. Clin Infect Dis 2009; 49:1821–7.
- Bouza E, Muñoz P, López-Rodríguez J, et al. A needleless closed system device (CLAVE) protects from intravascular catheter tip and hub colonization: A prospective randomized study. J Hosp Infect 2003; 54:279–87.
- 12. Casey AL, Worthington T, Lambert PA, Quinn D, Faroqui MH, Elliott TS. A randomized, prospective clinical trial to assess the potential infection risk associated with the PosiFlow needleless connector. J Hosp Infect **2003**; 54:288–93.
- Sawyer M, Weeks K, Goeschel CA, et al. Using evidence, rigorous measurement, and collaboration to eliminate central catheterassociated bloodstream infections. Crit Care Med 2010; 38:S292–8.
- Miller MR, Griswold M, HarrisII JM, et al. Decreasing PICU catheter-associated bloodstream infections: NACHRI's quality transformation efforts. Pediatrics 2010; 125:206–13.
- Sannoh S, Clones B, Munoz J, Montecalvo M, Parvez B. A multimodal approach to central venous catheter hub care can decrease catheter-related bloodstream infection. Am J Infect Control 2010; 38:424–9.
- Maki DG. In vitro studies of a novel antimicrobial luer-activated needleless connector for prevention of catheter-related bloodstream infection. Clin Infect Dis 2010; 50:1580–7.
- Casey AL, Mermel LA, Nightingale P, Elliott TSJ. Antimicrobial central venous catheters in adults: A systematic review and metaanalysis. Lancet Infect Dis 2008; 8:763–76.
- Timsit JF, Schwebel C, Bouadma L, et al. Chlorhexidine-impregnated sponges and less frequent dressing changes for prevention of catheter-related infections in critically ill adults: A randomized controlled trial. JAMA 2009; 301:1231–41.
- Venkatesh M, Rong L, Raad I, Versalovic J. Novel synergistic antibiofilm combinations for salvage of infected catheters. J Med Microbiol 2009; 58:936–44.