

CME Blood Pressure Monitoring for the Anesthesiologist: A Practical Review

Karsten Bartels, MD,* Stephen A. Esper, MD, MBA,† and Robert H. Thiele, MD‡

Periodic, quantitative measurement of blood pressure (BP) in humans, predating the era of evidence-based medicine by over a century, is a component of the American Society of Anesthesiologists standards for basic anesthetic monitoring and is a staple of anesthetic management worldwide. Adherence to traditional BP parameters complicates the ability of investigators to determine whether particular BP ranges confer any clinical benefits. The BP waveform is a complex amalgamation of both antegrade and retrograde (reflected) pressure waves and is affected by vascular compliance, distance from the left ventricle, and the 3D structure of the vascular tree. Although oscillometry is the standard method of measuring BP semicontinuously in anesthetized patients and is the primary form of measurement in >80% of general anesthetics, major shortcomings of oscillometry are its poor performance at the extremes and its lack of information concerning BP waveform. Although arterial catheterization remains the gold standard for accurate BP measurement, 2 classes of devices have been developed to noninvasively measure the BP waveform continuously, including tonometric and volume clamp devices. Described in terms of a feedback loop, control of BP requires measurement, an algorithm (usually human), and an intervention. This narrative review article discusses the details of BP measurement and the advantages and disadvantages of both noninvasive and invasive monitoring, as well as the principles and algorithms associated with each technique. (Anesth Analg 2016;122:1866–79)

The purpose of this narrative review¹ is to discuss the details of blood pressure (BP) measurement, including its history, the advantages and disadvantages of available noninvasive and invasive monitoring, the principles and algorithms associated with each technique, the engineering theories that describe how BP can be optimally managed, as well as available clinical data on BP management.

HISTORICAL INTRODUCTION

Qualitative analysis of the systemic arterial pulse BP waveform has been in practice for millennia, having been used by the Egyptians, Greeks, and Han Chinese.^{2,3} Construction of a physiologic model that matched anatomic observations with clinical experience would not occur for another 15 centuries, beginning with Harvey's demonstration that arteries and veins exist in series. This discovery was followed by independent work by both Borelli and Hales examining the concept of arterial elasticity with dampened flow in the periphery.⁴

In 1733, Hales performed the first known direct measurement of arterial BP in a white mare. He inserted a brass tube into the horse's artery, affixing it to an approximately 3.5-m glass tube with the flexible trachea of a goose. The height of the column of blood reportedly rose 2.51 m (approximately 185 mm Hg). Hales noted BP oscillations in the column, corresponding

with both ventricular heart rate and respiratory rate.^{2,5} Over 100 years later, J. Faivre, a French surgeon, performed the first known direct measurement of arterial BP in a human, connecting the superficial femoral artery of an amputee directly to a manometer.^{2,5} Although historically significant, Faivre's technique was clinically impractical at the time.

Because it required invasive arterial cannulation, direct measurement of arterial BP was typically performed only in the context of animal research until the mid-19th century. One hundred years after Hales' experiment, Jules Herrison produced the first noninvasive BP monitor, consisting of a mercury-filled bulb attached to a glass column. Application of the bulb to the radial artery at increasing pressures would eventually produce cessation of pulsatility in the fluid column, the point at which systolic pressure is estimated.⁴ In 1855, Karl Vierordt developed the first prototype for the sphygmomanometer, using a lever and weight system attached to a recording arm to obliterate the pulse.² In 1860, Jules Maray developed a more accurate variation of this device. Although too cumbersome for clinical use, it was a major advance for BP research because investigators could accurately measure systolic and diastolic BP.²

In 1901, Theodore Janeway first published the suggestion that auscultation could be used to quantify arterial BP measurements: "... certain experiments in a number of cases concerning the pressure tone and murmur in the brachial, to be described later, show that the production of the tone always occurs at a lower pressure than the point in question (disappearance of secondary waves) It is to be hoped that some more satisfactory method for estimating mean arterial pressure may yet be devised."⁴ The auscultatory technique, as used today, would be reported 4 years later by a Russian surgeon, Nicolai Korotkov.^{2,5} Used for decades, the Korotkov technique eventually adapted to the oscillometric technique, which allowed for the automated, noninvasive measurement of BP.

DOES BP MATTER?

Implicit in the inclusion of semiautomated BP measurements as a standard in the American Society of Anesthesiologists

From the *Department of Anesthesiology, University of Colorado Hospital, Aurora, Colorado; †Department of Anesthesiology, Cardiovascular and Thoracic Division, University of Pittsburgh, Pittsburgh, Pennsylvania; and Departments of ‡Anesthesiology and §Biomedical Engineering, University of Virginia, Charlottesville, Virginia.

Accepted for publication February 24, 2016.

Funding: None.

The authors declare no conflicts of interest.

Reprints will not be available from the authors.

Address correspondence to Robert H. Thiele, MD, Department of Anesthesiology, University of Virginia, P.O. Box 800710, Charlottesville, VA 22908. Address e-mail to rht7w@virginia.edu.

Copyright © 2016 International Anesthesia Research Society
DOI: 10.1213/ANE.0000000000001340

monitoring guidelines are the assumptions that BP is physiologically relevant and that deviation outside a particular range has clinical consequences.⁴ Systemic BP provides the energy source required to transport oxygen and nutrients to cells and tissues, and it is worth asking whether there is a relationship between BP and cardiac output and/or organ-specific blood flow. Although the physiology of the cardiovascular system is complex, especially as it relates to autoregulation and what appears to be a blood flow hierarchy at the level of individual organs,^{6,7} it is worth pointing out that in human beings undergoing general anesthesia, there is almost no relationship between mean arterial BP and delivery of oxygen (DO_2), presumably because of widely varying systemic vascular resistances⁸ (Fig. 1). Furthermore, in organs that autoregulate (e.g., the brain), BP and organ-specific blood flow are clearly uncorrelated over a range of pressure.⁹

To our knowledge, there are no known prospective, randomized controlled trials comparing 2 different defined ranges of BPs in patients undergoing surgery and general anesthesia. There have been such studies in the context of BP management during cardiopulmonary bypass, which have indicated fewer neurologic and cardiac complications in patients managed with higher versus lower mean arterial BP goals (80–100 mm Hg versus 50–60 mm Hg¹⁰ and 60–70 mm Hg versus 80–90 mm Hg¹¹). In the critical care literature, the Sepsis and Mean Arterial Pressure (SEPSISPAM) trial randomly assigned septic patients to mean arterial pressure (MAP) 65 to 70 mm Hg (low-target group) and MAP 80 to 85 mm Hg (high-target group) and found no difference in the primary outcome (28-day mortality).¹²

BP has also been included as a component in some perioperative and critical care hemodynamic management protocols, most notably as a component of Rivers' Early Goal-Directed Therapy sepsis protocol, where a MAP ≥ 65 mm Hg was associated with a substantial reduction in mortality.¹³ This observation was subsequently refuted by 3 larger, multicenter studies.^{14–16} BP targets have been sparingly used as a part of intraoperative goal-directed therapy algorithms. In 2 meta-analyses of perioperative goal-directed therapy totaling 61 studies,^{17,18} only 2 studies were identified, which used BP as a hemodynamic end point—one of which was negative.¹⁹

This is not to say that BP holds no clinical relevance. Indeed, because of its long history and relative ease of measurement, randomly assigning patients to what most would consider to be “low” BP may be a real challenge, especially considering varying patient populations. Thus, the question of how much hypotension is acceptable may never be answered in a prospective fashion, and large, complex trials may be required to begin answering this question in a rigorous fashion. In addition, there may be clinically significant differences related to how BP is maintained within a particular range (e.g., administration of fluid versus vasoactive agents). Vasoactive agents in particular have differing physiologic effects with, in some cases, significant clinical implications. A comprehensive review of fluid management strategies and detailed comparison of available




Figure 1. Correlation of O_2 delivery (DO_2) with mean arterial pressure (MAP) in the perioperative period in patients undergoing abdominal aortic surgery. Figure used with permission from Springer. Reinhart K. Zum Monitoring des Sauerstofftransportsystems. *Der Anaesthesist* 1988;37:1–9. © Springer - Verlag 1988.⁸

vasoactive agents are beyond the scope of this review but are available elsewhere.^{6,7,20,21}

A retrospective analysis of 24,120 patients undergoing noncardiac anesthetics, the “Triple Low” study, investigated the impact of MAP, Bispectral Index score, and minimum alveolar concentration on 30-day mortality. It found that the relative risk of death for patients who had isolated low BP (defined as MAP < 81.7 mm Hg) was 0.729 (95% confidence interval [CI], 0.342–1.558).²² However, in another large, single-center retrospective study of 33,330 noncardiac surgeries, a more conservative threshold to define hypotension (MAP < 55 mmHg), for even brief periods of time (1–5 minutes), was associated with increased relative risks of acute kidney injury (1.18 [95% CI, 1.06–1.31]) and myocardial injury (1.3 [95% CI, 1.06–1.5]).²³

PHYSIOLOGIC BASIS OF SYSTEMIC ARTERIAL PRESSURE WAVEFORM

In subjects with normal body physiology, BP waveforms are produced by left ventricular contraction. Although the left ventricular pressure waveform may be relatively simple, the resultant peripheral pressure waveform is a function of distance from the aortic valve, compliance of the systemic arterial blood vessels, and geometric considerations (in particular, arterial diameter and arterial branch points).³

In the absence of disease, left ventricular pressure is transmitted directly to the ascending aorta during systole. Most of the ejected blood will pass through the aorta, which primarily serves as a conduit. Because the aorta is compliant, it expands as pressure rises, storing blood and ventricular energy, in the form of pressure, to be used during diastole. As systole comes to an end, ventricular pressure falls rapidly, and forward flow through the aortic valve diminishes. When the aortic valve closes and diastole begins, the left ventricle

⁴Available at: <http://www.asahq.org/~media/Sites/ASAHQ/Files/Public/Resources/standards-guidelines/standards-for-basic-anesthetic-monitoring.pdf>. Accessed February 1, 2016.

and ascending aorta are no longer hydraulically connected. Aortic pressure then becomes a function of the interaction between the compliant aorta, which expands and contracts, and the downstream arterial beds, which resist flow. This simplified model of aortic blood flow is known as the Windkessel model, developed by Otto Frank.^{24–26} In what is referred to as the “two-element Windkessel model,” the aorta behaves as a capacitor and the tissue beds as a resistor. Models of increasing complexity, which include inductance and aortic input impedance, have been developed.²⁷

Although the Windkessel model was a major advance in our understanding of BP and hemodynamics, it is an oversimplification of reality and has shortcomings. One of the most relevant assumptions of this model is that the aorta is modeled as a single, infinitely long tube of consistent diameter and compliance. In reality, the aorta is finite, gradually branching off until becoming the left and right common iliac arteries. These major arteries lead to arterioles and the capillary bed, which, collectively, have a much larger cross-sectional area than the aorta and completely different mechanical properties.

Changes in aortic diameter (leading to changes in resistance) and compliance affect impedance or resistance to pulsatile flow. These impedance changes are mainly responsible for reflected pressure waves traveling backward, from the periphery toward the left ventricle.^{28,29} Thus, the shape of the systemic arterial pressure waveform at any point along the arterial tree is a sum of the forward traveling (incident) wave, as well as any reflected waves (Fig. 2). Different arteries have different shapes, and these shapes change over time. As individuals age, and arterial compliance is diminished, the summation of forward and backward travelling waves leads to a gradual increase in pulse pressure in the direction of the periphery.

BP MONITORING TECHNOLOGY

Sphygmomanometry and Oscillometry

Sphygmomanometry

An understanding of the pulsatile nature of the systemic arterial blood stream, combined with previous efforts at estimating BP using primitive devices, allowed Korotkov to develop the auscultatory technique for BP measurement. Korotkov’s major intellectual contribution to BP measurement was the recognition that the application of external

pressure exceeding systolic BP would completely eliminate blood flow and that as this external pressure was gradually reduced, blood flow would return in a predictable manner.

The Korotkov technique required the use of both a sphygmomanometer (inflatable cuff with pressure gauge) and a stethoscope to link pressure changes with return of blood flow. A major disadvantage of this technique is that it requires enough blood flow to generate the turbulence needed to produce Korotkov sounds. In instances in which blood flow to an extremity may be compromised (e.g., central redistribution in the setting of shock), this technique may fail secondary to the inability to detect Korotkov sounds.³⁰ Furthermore, sphygmomanometry also requires that cuff inflation lead to arterial occlusion. In extremely large patients, or in those with excessively calcified (and rigid) arteries, this may not be possible. Despite these limitations, sphygmomanometry is generally accurate in relatively healthy, normal-sized patients.³¹

Unfortunately, sphygmomanometry is time consuming and labor intensive. Because it does not lend itself well to automation, semicontinuous measurement and use during anesthesia are impractical. It would not be until the development of oscillometry, a mechanical extension of sphygmomanometry, that automated, semicontinuous measurement of BP would be possible (Fig. 3).

Oscillometry

The concept of oscillometry was first developed by Marey³² in mid-19th century. Sixteen years after first demonstrating the arterial “sphygmograph” in 1860, Marey conducted a series of experiments in which the forearm was placed in a water-filled chamber to which counterpressure was applied. This work formed the foundation for both the recognition that the magnitude of these pulsations could be used to approximate BP (oscillometric technique)^{33–35} and the appreciation that a counterpressure waveform designed to perfectly eliminate the pulsatility of distal tissue bed could accurately reproduce the BP waveform (volume clamp technique).³⁶

Unlike the auscultatory technique, in which the return of blood flow is detected audibly, the oscillometric technique measures the changes in cuff pressure that occur when blood flow returns during deflation. As demonstrated by Yelderman and Ream,³³ the amplitude of pressure changes detected in the cuff is near-maximal when the cuff is inflated at MAP. Systolic and diastolic pressures are estimated using proprietary algorithms unique to each manufacturer and should not be considered as accurate as MAP.^{33,35,37}

Because the physical principles on which they are based are similar, the oscillometric technique suffers some of the same disadvantages as the auscultatory technique, including poor performance in larger patients and patients with calcified arteries, as well as in those with reduced blood flow. Because the oscillometric technique uses an inflatable cuff to both stop blood flow and detect return, proper cuff sizing is particularly important. In general, the cuff dimensions, indexed to arm circumference, should be 80% in length and 40% in width.³⁷ As might be expected, oscillometric measurements are site dependent.³⁸ Oscillometry is particularly challenging in neonates because of their small size (cuff sizing is crucially important) and artifacts associated with crying and moving. Furthermore, these devices may not be accurate in measuring BP in infants of low birth weight who are hypotensive.³⁹




Figure 2. Representation of forward and reflected pressure wave travel and the influence of their timing and overlap on recorded aortic and peripheral pressure waves. T_{sh} = time to shoulder. Figure used with permission from Elsevier. Briet M, Boutouyrie P, Laurent S, London GM. Arterial stiffness and pulse pressure in CKD and ESRD. *Kidney International* 2012;82:388–400. © Elsevier 2012.¹¹⁷




Figure 3. Cuff pressure with superimposed Korotkoff sounds and amplified cuff pressure oscillations. S_0 is the point where cuff-pressure oscillations start to increase. A_s is the amplitude corresponding to auscultatory systolic pressure and A_d is the amplitude corresponding to auscultatory diastolic pressure. A_m is the maximum oscillation amplitude, which signals mean pressure. Figure used with permission from John Wiley & Sons, Inc., New York, NY. Geddes, LA. Cardiovascular Devices And Their Applications. Chapter 3, "Blood Pressure: Noninvasive Measurement", p 72. © John Wiley & Sons, Inc. 1984.¹¹⁸

Although oscillometry is typically performed in an automated fashion, it can be performed manually using a cuff and aneroid manometer—as cuff pressure drops below systolic BP, the needle will begin to oscillate with each pulsation, increasing in amplitude until MAP is approached and then diminishing toward diastole.

Accuracy

One of the largest comparisons of invasive to noninvasive BP measurements was made by Wax et al.,⁴⁰ who retrospectively compared BP from automated oscillometric readings with arterial BP readings derived from intravascular catheters in >15,000 noncardiac anesthetics. They found the SD of the difference between the 2 to be approximately ± 12.5 mm Hg at an MAP of 75 mm Hg (which was the mode of the distribution). At the extremes (hypertension or hypotension), the SD worsened substantially, further weakening the utility of this ubiquitous (and American Society of Anesthesiologists standard) class of monitoring devices. In a meta-analysis of studies comparing commercially available noninvasive arterial BP measurements with invasive measurements, the overall random-effects pooled SD was 8.4 mm Hg for MAP for all devices. Of note, this pooled measure did not specifically isolate oscillometric devices.⁴¹ A collection of 7 prospective studies comparing the oscillometric technique with invasive measurements of MAP is presented in Table 1.^{33,42–47}

Tonometry

In search of a noninvasive mechanism by which BP waveforms (as opposed to numerical estimates) could be measured, O'Rourke et al.⁴⁸ developed the field of arterial tonometry. Tonometers "measure tone" by placing a sensor directly over an artery, just as an anesthesia provider might do when palpating the groin in search of a femoral pulse. Tonometers, such as the Sphygmocor (AtCor, Sydney,

Australia), quantify and display this tactile information. The Sphygmocor device has been validated in >1600 patients (including >15,000 individual measurements), but because it requires unrestricted access to the wrist to apply a penlike tonometric sensor, it is primarily used for research.⁴⁸

In an effort to develop a more user-friendly and automated tonometer, Tensys Medical (Irvine, CA) affixed a tonometer to a locking clamp that stabilizes the tonometer over the radial artery in the wrist—the T-line®. Once affixed over the radial artery, it can periodically locate the ideal measuring point and adjust the tonometric pressure sensor. This feature allows the device to function with minimal operator intervention and is analogous to the development of the oscillometer (which made automated cuff-based BP measurements possible).

Several studies have compared invasive arterial measurements of mean arterial BP with estimates from a tonometer during general anesthesia and surgery. The median SD of the difference between the 2 is 6.3 mm Hg (range 4.8–12.6 mm Hg).^{49–53} Assuming no bias, the limits of agreement are approximately ± 12.4 mm Hg (range ± 9.4 to ± 24.7 mm Hg). In a meta-analysis of studies comparing commercially available noninvasive arterial BP measurements with invasive measurements, the overall random-effects pooled SD was 5.7 mm Hg for MAP using the T-line, slightly better than the 8.4 mm Hg reported for all noninvasive devices overall, suggesting that tonometric devices outperform oscillometric techniques.⁴¹

Pulse Wave Velocity

Pulse wave velocity (V) in an elastic tube is a function of elasticity (E), diameter (D), wall thickness (h), and fluid density (ρ).^{54,55} Mathematically this is known as the Moens-Korteweg equation⁴:

$$V = k(Eh / \rho D)^{1/2}$$

Table 1. Mean Arterial Blood Pressure Measured by Intraarterial Versus Oscillometric Technique

Study	Population	Device	Bias, mm Hg	SD (95% CI), mm Hg
Heard et al. (2000) ⁴²	Intensive care unit patients requiring intraarterial monitoring (n = 28)	Dinamap, Critikon, Tampa Bay, FL	-1.3	6.3 (-17.7 to 13.7)
Yelderman and Ream (1979) ³³	Patients undergoing open heart surgery (n = 19)	12 or 13-cm cuff applied above brachial artery and calibrated using mercury manometer	1.4	6.22 (-4.1 to 11.7)
Borow and Newburger (1982) ⁴³	Patients undergoing cardiac catheterization (n = 30)	Dinamap	-3.5	4.9 (-13.1 to 6.1)
Loubser (1986) ⁴⁴	Hypertensive post-carotid endarterectomy patients (n = 30)	Dinamap	0.26	Not reported
Gravlee and Brockschmidt (1990) ⁴⁵	Adults undergoing cardiac surgery (n = 38)	Dinamap	6	8 (-9.68 to 21.7)
Gorback et al. (1991) ⁴⁶	Patients requiring intraarterial monitoring (n = 32)	Dinamap	-1.8	9.7 (-20.8 to 17.2)
Manolio et al. (1988) ⁴⁷	Patients undergoing vascular or other major surgery (n = 14)	Dinamap	2.9	1.7 (-0.4 to 6.2)

Comparison studies of intraarterial blood pressure versus blood pressure measured using oscillometric technology. If either 95% CI or SD were not reported, the missing values were calculated using the formula $CI = Bias \pm 1.96 \times SD$. Bias was recalculated from the published data as necessary to standardize: $Bias = MAP_{oscillometric} - MAP_{intraarterial}$ for all studies reported. CI = confidence interval; MAP = mean arterial blood pressure.

In theory, changes in vascular tone should lead to changes in pulse wave velocity, calculated as the time between an R wave on the electrocardiogram and the appearance of a peripheral pulse waveform (detected by a photoplethysmogram [PPG] or an invasive arterial pressure tracing) divided by the distance between the heart and the site of measurement. In situations in which cardiac output is relatively constant, changes in vascular tone will lead to changes in BP. Thus, in theory, pulse wave velocity (or pulse transit time [PTT]) may give some indication of BP.

In one early study, PPG-based PTT was successfully used to estimate BP in patients undergoing general anesthesia⁵⁶; however, a larger, multicenter study conducted 5 years later refuted these findings.⁵⁷ There are several reasons why PTT alone will not likely lead the clinician to absolute measures of BP. First, several of the individual components of the Moens-Korteweg equation (E, h, D) are not typically known for individuals. Second, the Moens-Korteweg equation neglects the viscous properties of blood-filled arteries, whose behavior can best be described as viscoelastic.⁵⁸ Third, cardiac output is not stable in individuals and is widely variant between them. For these reasons, PTT is more likely to achieve success as a trend monitor valid over short time periods.

Indeed, PPG-PTT was recently used to detect changes in BP during induction of both general and neuraxial anesthesia.^{59,60} Although more studies are needed, the rapid response of these noninvasive monitors is encouraging and suggests they may be useful for alerting the clinician to probable hemodynamic changes before they might be detected using automated oscillometric techniques.

Volume Clamp Technique

As described earlier, Marey's work applying counterpressure to a submerged forearm formed the foundation for the volume clamp technique, which was formally developed by Peñáz,³⁶ a Czech physiologist.⁶¹ Peñáz' major advance was to connect a pressure cuff (on the finger) to a distally placed PPG recorder using a feedback loop, which produced ultrafast adjustments in bladder pressure in response to changes in the PPG signal. To maintain a perfectly flat PPG

waveform, increasing cuff pressure is applied in response to increasing peripheral artery pressure. Conversely, decreasing cuff pressure is applied in response to decreasing peripheral artery pressure. The cuff pressure waveform required to maintain a flat PPG waveform closely approximates the systemic arterial pressure waveform.⁶²

Although a control loop modifying finger bladder pressure in response to changes in a PPG tracing can estimate the shape of a peripheral arterial pressure waveform (i.e., measuring the qualitative aspects of the systemic arterial BP tracing), more work is needed to provide quantitative measures of systolic, diastolic, and MAPs. Again, previously established principles are helpful. Yelderman and Ream³³ demonstrated that the amplitude of pressure changes detected in a BP cuff is near-maximal when the cuff is inflated at or near MAP. By measuring the amplitude of the PPG waveform distal to a finger cuff while applying a constant pressure to the finger cuff, then changing the pressure in either direction such that the amplitude of the PPG waveform is maximal, MAP can be estimated.

Because MAP changes over time, volume clamp BP monitors have to periodically recalibrate themselves to reidentify MAP. When an external pressure cuff reaches MAP, the arterial wall stress is minimized, and the artery beneath the cuff is said to exist in an "unloaded" or "unstretched" state. As external pressure decreases, the forces intrinsic to the arterial wall take over the pressure "load," gradually increasing the arterial wall stress required to maintain vessel shape.^{63,64} To keep the underlying artery in an unstretched state, volume clamp devices modify cuff pressure periodically, analyzing the PPG to identify the pressure needed to produce an unstretched artery (i.e., the point at which PPG amplitude is maximal). Wesseling combined the volume clamp technique with Peñáz' feedback loop, producing the Finapres™ (Finapres Medical Systems BV, Amsterdam, The Netherlands) device.^{63,65,66}

Modern volume clamp BP monitors use slightly different techniques to achieve vascular unloading. The ClearSight™ (Edwards Lifesciences, Irvine, CA) device, which was developed as the Nexfin® (BMEYE, Amsterdam, The Netherlands) device, is a direct extension of Wesseling's Finapres device and the "Physiocal" (physiologic calibration) algorithm to




Figure 4. Finger blood pressure measurement using photoplethysmography using the volume-clamp and physioal approaches. Figure used with permission from Springer. Bartels K, Thiele RH. Advances in photoplethysmography: beyond arterial oxygen saturation. *Can J of Anaesth* 2015;62:1313–28. © Springer 2015.⁶⁸

achieve vascular unloading. The Physioal strategy induces periodic, stepwise bladder pressure changes and simultaneously reanalyzes the PPG waveforms. If alternate pressure settings result in larger PPG amplitudes, the feedback set point is adjusted to ensure a maximally unloaded state of the finger artery⁶³ (Fig. 4).

The **Continuous Noninvasive Arterial Pressure (CNAP®;** CNSystems, Graz, Austria) device uses the “Vasomotor Elimination and Reconstructed Identification of the Initial set point” (**VERIFI**) algorithm to unload peripheral arteries, coupled with an upper arm oscillometric BP monitor for calibration. The CNAP is designed to account for the fact that **no feedback mechanism can be completely instantaneous** (thus introducing error because of the time lag between the cuff and the PPG signal detector).^b To generate an initial BP measurement and identify the “set point,” the VERIFI algorithm **relies on a brachial artery cuff–Noninvasive Blood Pressure (NIBP)**. The volume clamp (finger bladder) is still used to approximate the BP waveform by maintaining constant blood volume as a function of time, $V(t)$. However, the VERIFI algorithm uses a different set point calibration strategy than the Physioal solution.⁶⁷

Like musical notes, physiologic waveforms (e.g., BP) can be described as the sum of a series of sine waves, each with its own **amplitude, frequency, and phase shift**. The dominant waveforms, whose frequencies center around multiples of the fundamental frequency, are known as “**harmonics**.”⁷³ The VERIFI algorithm decomposes the volume signal, $V(t)$, into frequency components using frequency domain analysis (e.g., **Fourier transformation**). Pulsatile volume, $V_{AC}(t)$, low frequency volume, $V_{LF}(t)$, and very low frequency volume, $V_{VLF}(t)$, are isolated. Because volume changes are a function of pressure changes, volume and pressure can be related to

each other by the degree of amplification, also known as “**gain**” (G , analogous to the resistance in Ohm’s law). Gain is assumed to be zero at very low frequencies, but at the “**pulsatile**” and low frequencies (G_{AC} , G_{LV} , respectively) can be determined by comparing various fixed cuff pressures to the signal amplitude and developing a transfer function (T) relating the 2 (Fig. 5).⁶⁷ For a comparison of the features of the Nexfin/ClearSight and CNAP devices, see **Table 2**.⁶⁸

Clinical Considerations and Accuracy

As with sphygmomanometry and oscillometry, **volume clamp devices** require peripheral **perfusion** and blood flow to estimate BP. In theory, just as the **pulse oximetry waveform** (which is **derived from both red and infrared PPG signals**) is affected by extreme **vasoconstriction, temperature changes, and other physiologic disturbances**, the **same** should be true of **volume clamp estimates of BP**.

The impact of **vasomotor tone** on this class of devices has been studied. An early analysis of the Finapres device during surgery found that half of the measurements were conducted in a state of vasoconstriction (defined as a PPG amplitude reduction of at least 50% [PPG amplitude is generally considered an acceptable surrogate measure for vasomotor tone⁶⁹]). This small study, in which 378 paired measurements were made in 6 subjects, suggested that the **discrepancy between volume clamp and oscillometric devices was dependent on peripheral perfusion**.⁷⁰ Interestingly, a direct comparison between the Finapres and the invasive measurements suggests that the **thumb is the optimal site for finger arterial BP measurements**, presumably because of the **larger arterial cross-sectional area and better perfusion**.⁷¹

In a meta-analysis of studies **comparing commercially available noninvasive arterial BP measurements with invasive measurements**, the overall random-effects pooled **SD was 9.4 mm Hg for MAP using the CNAP, slightly worse than the 8.4 mm Hg reported for all noninvasive devices overall**.⁴¹ This study

^bAvailable at: <http://www.cnsystems.com/Innovation/Cnap-Technology>. Accessed February 1, 2016.




Figure 5. Decomposition of a standard pulse oximetry signal into three bands: a very low frequency band (0-0.1 Hz), a low frequency band (0.1-0.5 Hz), and a pulsatile band (0.5-2 Hz). Figure used with permission from Springer. Bartels K, Thiele RH. Advances in photoplethysmography: beyond arterial oxygen saturation. *Can J Anaesth* 2015;62:1313–28. © Springer 2015.⁶⁸

did not specifically analyze data for the Nexfin or ClearSight devices. Because a true gold standard is readily available (direct arterial catheterization), volume clamp devices will always be less accurate than the gold standard. The important clinical question is how well the volume clamp devices agree with direct catheterization, and how this compares with other noninvasive devices (e.g., oscillometry). However, as examined in an aforementioned study comparing oscillometry and direct arterial catheterizations, as the true BP deviated from the mode, oscillometric performance deteriorated, particularly at the extremes (hypertension, hypotension) when accurate measurements matter the most.⁴⁰

There have been multiple comparisons of volume clamp devices with direct arterial catheterization. Because

this body of work spans several decades, not all of them use the same statistical techniques. More modern studies use the limits-of-agreement approach. Overall, these data suggest performance that is at least comparable to oscillometric techniques, although there are no large, retrospective comparisons available to validate the prospective observations in the real-world clinical environment (Table 3).^{72–96}

Arterial Catheterization

Background

In addition to giving the clinician a near-instantaneous measurement of systolic, diastolic, and MAPs, arterial catheterization also allows for convenient, periodic blood sampling

Table 2. Comparison of Nexfin to CNAP Volume Clamp Devices

Characteristic	CNAP	Nexfin/ClearSight
Fingers used	2	1–2
Algorithm to continuously recalibrate blood pressure (account for changes in vasomotor tone)	VERIFI	Physiocal
Algorithm for stroke volume	Pulse contour analysis using CNCO algorithm and biometric or manual calibration	Pulse contour analysis and biometric calibration

Comparison of Nexfin to CNAP Volume Clamp Devices for the continuous measurement of both blood pressure and stroke volume. Note that while both devices rely on the volume clamp technique, they employ different algorithms to consistently measure blood pressure in the setting of changing vascular tone and apply different arterial waveform analysis algorithms for the estimation of stroke volume.

CNAP = continuous noninvasive arterial pressure; CNCO = continuous noninvasive cardiac output; VERIFI = Vasomotor Elimination and Reconstructed Identification of the Initial set point.

(e.g., for the measurement of arterial blood gases). Analysis of >12,000 surgical cases at an Academic Medical Center in the United States revealed that 16% of patients were monitored with an arterial catheter.⁹⁷ By 1990, >8 million had been placed.⁹⁸

A variety of algorithms designed to measure stroke volume (SV) and fluid responsiveness (e.g., pulse pressure variation, SV variation) based on sophisticated analysis of the arterial waveform have been developed.⁹⁹ Until the appearance of the volume clamp technique devices on the market, noninvasive estimates of SV, pulse pressure variation, and SV variation required placement of an intraarterial catheter. Now, with the availability of the ClearSight and the CNAP, the same algorithms can be applied to an arterial pressure waveform derived from the volume clamp technique. At present, there are insufficient clinical data to comment on the relative accuracy of noninvasive SV derived from directly or indirectly measured arterial pressure waveforms.

That said, volume clamp devices, which seem to outperform automated oscillometric techniques in terms of accuracy and measurement frequency, may one day change the anesthesiologist's decision-making process with regard to BP monitoring.¹⁰⁰ This is particularly important, as the complications associated with percutaneous cannulation of a systemic artery, including infection, pseudoaneurysm formation, nerve injury, permanent arterial occlusion, or hematoma formation, are increasingly appreciated.^{101–103}

Arterial BP is most commonly measured at the radial, brachial (or axillary), and femoral sites, with the radial being most common.¹⁰³ As described earlier, the BP waveform is a function of both incident and reflected waves and is dependent on arterial compliance, branch points, and distance from the left ventricle. Thus, different arteries produce different BP tracings within the same individual.

Procedure

Arterial cannulation can be performed using a variety of techniques. The simplest strategy is to insert a 20-gauge (or smaller) catheter directly into the artery and "thread"

the catheter off a needle and into the artery. Some clinicians prefer to "transfix" the artery with a needle and catheter, remove the needle, and retract the catheter backward until it resides in the vessel lumen, after which a guidewire can be inserted through the catheter and into the artery. For particularly challenging patients, the Seldinger technique can be used, in which a needle is inserted into the vessel, a guidewire is threaded through the needle and into the vessel, the needle is removed, and the catheter is threaded into the vessel (over the wire). Identification of the artery may be facilitated by using ultrasound guidance.¹⁰⁴ A meta-analysis of 7 trials including 482 patients suggested that ultrasound-guided radial artery cannulation increased the first-attempt success rate, as well as decreasing the number of attempts, time to success, and rate of hematoma formation.¹⁰⁵

Some clinicians use the "Allen test" to evaluate collateral flow to the hand before cannulating the radial artery. This test is performed by occluding arterial inflow into the hand (by compressing both the radial and the ulnar arteries simultaneously) while asking the patient to create a fist, thereby exsanguinating the palm. After opening the hand and releasing the pressure over the ulnar artery (but maintaining pressure over the radial artery), a qualitative sense of the adequacy of ulnar flow can be gained by observing how quickly the palm returns to color. Although the Allen test seems logical, it is not clear that reason and reality have aligned, because large analyses including several thousand patients have questioned the predictive ability of the Allen test.^{106,107} Furthermore, it is now appreciated that the radial artery can often be harvested for coronary bypass surgery, despite an abnormal Allen test, and this has prompted some investigators to look for alternative testing modalities.^{108,109}

Equipment

Intraarterial pressure is measured using a strain gauge, the essential component of which is a thin metal wire whose resistive properties change as it is stressed. By building strain gauges into a Wheatstone bridge circuit, changes in strain can be measured as electrical current through the bridge fluctuates. These changes in current can be translated into pressure changes by "zeroing" the transducer. Importantly, current intraarterial pressure monitors measure the pressure of a fluid-filled column that is in hydrostatic continuity with the blood stream. Thus, the physical attributes of the measuring system—in particular, the length, diameter, and stiffness of tubing—can potentially affect the measured BP waveform.

The fluid-filled transduction systems that are commonly used to transmit intravascular pressure changes to a Wheatstone bridge strain gauge can be described as "underdamped, second-order dynamic systems."¹¹⁰ Three parameters define the performance characteristics of these systems: mass (of fluid), elasticity (of tubing), and friction (primarily between the fluid and the tubing). Every fluid-filled tubing system has a natural frequency (f_n), that is, a frequency at which a pressure pulse will oscillate within the system, and a damping coefficient (ζ), a measure of how quickly an oscillating waveform will decay. Fluid-filled transduction systems are "underdamped" because a pressure pulse will oscillate at the natural frequency of the system before decay. Rigid, short, narrow diameter tubing more rapidly transmits pressure waveforms, effectively

Table 3. Mean Arterial Blood Pressure Measured by Volume Clamp (ClearSight/Nexfin or CNAP) Versus Intraarterial Technique

Study	Patient population	Device	Bias, mm Hg	SD (95% CI), mm Hg
Stover et al. (2009) ⁷²	Critically ill (n = 10)	Nexfin	2	8 (−13.7 to 17.7)
Biais et al. (2010) ⁷³	Major vascular surgery (n = 25)	CNAP	1.8	10.3 (−18.4 to 22)
Hofhuizen et al. (2010) ⁷⁴	Pediatric congenital heart surgery (n = 10)	Nexfin	0.19	2.7 (−5.1 to 5.5)
Jelezov et al. (2010) ⁷⁵	Elective abdominal surgery, cardiosurgery, or neurosurgery (n = 88)	CNAP	1.6	11 (−20 to 23.2)
Martina et al. (2010) ⁷⁶	Cardiopulmonary bypass (n = 18)	Nexfin	−1.3	6.5 (−14 to 11.4)
Schramm et al. (2011) ⁷⁷	Transfemoral aortic valve implantation with sedation (n = 29)	CNAP	−0.8	15 (−30.2 to 28.6)
Fischer et al. (2012) ⁷⁸	Post-cardiac surgery intensive care (n = 44)	Nexfin	4.6	6.5 (−8.1 to 17.3)
Garnier et al. (2012) ⁷⁹	Pediatric intensive care (n = 41)	Nexfin	−2.6	7.7 (−17.7 to 12.5)
Hahn et al. (2012) ⁸⁰	Elective surgery with general anesthesia (n = 100)	CNAP		
		50 patients with standard software	2.9	10.6 (−17.9 to 23.7)
		50 patients with modified software	3.1	9.5 (−15.4 to 21.6)
Ilies et al. (2012) ⁸¹	Surgery with general anesthesia (n = 85)	CNAP		
		Induction of anesthesia	10.2	11.3 (−10.8 to 32.3)
		Maintenance of anesthesia	4.3	10.4 (−16.1 to 24.7)
Jagadeesh et al. (2012) ⁸²	Cardiac surgical intensive care (n = 30)	CNAP	−0.04	2.05 (−4.1 to 4)
Martina et al. (2012) ⁸³	Cardiac surgery (n = 50)	Nexfin	2.2	6.4 (−10.3 to 14.7)
Broch et al. (2013) ⁸⁴	Cardiac surgery (n = 50)	<i>Nexfin versus femoral</i>		
		Pre-CPB	6.2	11.7 (−16.7 to 29.2)
		Post-CPB	9.7	11 (−11.9 to 31.3)
		<i>Nexfin versus radial</i>		
		Pre-CPB	10.9	13 (−14.6 to 36.5)
		Post-CPB	13.5	11.3 (−8.6 to 35.7)
Dewhurst et al. (2013) ⁸⁵	Pediatric surgery in the prone position (n = 20)	CNAP	−0.26	6.1 (−12.18 to 11.67)
Gayat et al. (2013) ⁸⁶	Surgery with general anesthesia (n = 52)	CNAP	8	14 (−18 to 35)
Hohn et al. (2013) ⁸⁷	Critically ill surgical (n = 25)	Nexfin	6	12 (−18 to 30)
Schramm et al. (2013) ⁸⁹	Elective transfemoral aortic valve implantation with monitored anesthesia care (n = 33)	CNAP	4	11.3 (−26.1 to 18.1)
Ameloot et al. (2014) ⁹⁰	Critically ill (n = 45)	Nexfin	2.3	12.4 (−22 to 26.6)
Ilies et al. (2015) ⁹¹	Cardiovascular surgical critical care (n = 104)	CNAP	6.1	7.6 (−10.1 to 22.4)
Kumar et al. (2015) ⁹²	Cardiac surgery (n = 60)	CNAP	0.02	5.68 (−11.12 to 11.17)
Tobias et al. (2014) ⁹³	Severely obese adolescents and young adults undergoing weight loss surgery (n = 18)	CNAP	0.6	8.6 (−16.2 to 17.4)
Vos et al. (2014) ⁹⁴	Elective general surgery (n = 110)	Nexfin	2	9 (−15 to 19)
Wagner et al. (2015) ⁹⁵	Intensive care unit (n = 38)	CNAP	1	9 (−16 to 19)
Weiss et al. (2014) ⁹⁶	Elective surgery (n = 31) 11 time points	Nexfin		
		T-1 (n = 31)	7	14 (−21 to 34)
		T0 (n = 31)	7	16 (−24 to 38)
		T1 (n = 31)	7	14 (−21 to 35)
		T2 (n = 31)	7	17 (−26 to 40)
		T3 (n = 31)	5	12 (−18 to 29)
		T4 (n = 31)	7	11 (−15 to 30)
		T5 (n = 31)	8	12 (−15 to 31)
		T6 (n = 30)	8	13 (−18 to 34)
		T7 (n = 26)	9	15 (−20 to 39)
		T8 (n = 26)	8	13 (−17 to 33)
		T9 (n = 26)	9	17 (−24 to 41)

CI = confidence interval; CNAP = Continuous Noninvasive Arterial Pressure; MAP = mean arterial blood pressure.

Comparison studies of intraarterial blood pressure versus blood pressure measured using volume clamp technology (ClearSight/Nexfin or CNAP). If either 95% CI or SD were not reported, the missing values were calculated using the formula $CI = Bias \pm 1.96 \times SD$. Bias was recalculated from the published data as necessary to standardize: $Bias = MAP_{\text{volume clamp}} - MAP_{\text{intraarterial}}$ for all studies reported.

increasing the natural frequency. Dampening is a function of energy loss, which occurs primarily because of friction. Air bubbles, tubing kinks, stopcocks, and other irregularities can alter the physical characteristics of the tubing system and distort the recorded waveform.

Depending on the complexity of the BP waveform, as many as 10 harmonics may be required to accurately

reproduce the BP.^{3,110} Because the human heart rate is approximately 1 to 2 Hz (60–120 beats/min), higher-order harmonics may approach 20 Hz in frequency. To faithfully reproduce these higher-order harmonics, a natural frequency as high as 40 Hz is needed. The natural frequency of systems used clinically can be considerably lower than this.¹¹¹




Figure 6. Flush test. The natural frequency (f_n) is the frequency at which the pressure pulse will oscillate within the system and can be calculated from: $f_n = v/\lambda$. The damping coefficient (ζ) is a measure of how quickly an oscillating waveform will decay and is dependent on the ratio of successive amplitude ratios (e.g. A_1/A_2).¹¹⁰

Thus, waveform distortion is a function of both f_n and ζ , as well as the complexity and component frequencies of the pressure waveform (with more complex, higher-frequency waveforms being more prone to distortion). Acceptable combinations of f_n and ζ , which lead to accurately reproduced waveforms, are referred to as having adequate dynamic response. To assess f_n and ζ , the “flush test” can be used. Rapid introduction and removal of a high-pressure pulse (known in engineering lexicon as a “square wave”) into the transducing system will lead to a characteristic, oscillatory waveform that reverberates at f_n and diminishes over time in accordance with ζ (Fig. 6).

Interpretation

Measurement of intraarterial BP is relatively straightforward if the aforementioned principles on which the measurement system is based are understood. One common mistake is to attribute high systolic BPs to “whip” or an under-responsive system. Although it is possible for an under-responsive system to augment pulse pressure (thus raising systolic pressure and lowering diastolic pressure), this theory can be tested by performing a flush test, measuring both f_n and ζ and plotting the performance characteristics of the transduction system on a dynamic response map. In reality, it is normal for pulse pressure to increase as the measurement site moves distally, particularly in patients with non-compliant vasculature. This widened pulse pressure should initially be assumed to be real and only attributed to suboptimal equipment after a determination of both f_n and ζ has been made.

The BP waveform can be mathematically decomposed into a series of sine waves, each described by their amplitude, frequency, and phase shift. This is accomplished using a computational technique called “fast Fourier transformation.” The fundamental frequency of the BP waveform is the heart rate (for a heart rate of 80 beats per minute, the fundamental frequency is 1.33 Hz or 1.33 beats per second). When “harmonic” waves are added to the fundamental wave, the true shape of the waveform is more accurately approximated.

Complications

Although pseudoaneurysm formation, nerve injury, permanent arterial occlusion, and hematoma formation are known complications of intraarterial catheterization, the risk of infection is increasingly appreciated.^{101–103} An analysis of over 200 published studies on infectious complications associated with intravascular devices

estimates an infection rate of 1.7 per 1000 catheter-days for arterial catheters, when compared with 0.5 per 1000 for peripheral IV catheters and 2.7 per 1000 for nontunneled central venous catheters.¹⁰² A large, factorial trial analyzing the impact of dressing changes and chlorhexidine-impregnated sponge use on both central venous and arterial catheters demonstrated a substantially lower rate of bloodstream infections (hazard ratio, 0.39; 95% CI, 0.17–0.93), suggesting that these devices should be placed on all arterial catheters.¹¹²

FEEDBACK LOOPS, GAIN, AND BP CONTROL

Assuming that anesthesiologists desire to maintain BP within some prespecified range, a feedback loop must be established among the patient, the BP monitor, and the anesthesiologist. Feedback loops are well described in the engineering literature.¹¹³ A rudimentary feedback loop consists of a measurable variable (e.g., temperature), a method of measurement (e.g., thermometer), and a response (e.g., heating or cooling). The “set point” is the ideal value for the measured variable (e.g., 37°C). In the simplest feedback loop, the response is linearly related to the difference between the measured process variable and the set point. The magnitude of the response, indexed to this difference (known as an error value), is known as gain. This type of system is often referred to as a proportional–integral–derivative controller.

In real life, there is often a time delay associated with the detection of an abnormality both in the measured variable and in the response (known as response time). In a fairly stable system with minimal response time, large amounts of gain can exert substantial control over the measured variable. However, if the response time is long (either because there is measurement delay or it takes time to affect the measured variable [e.g., application of heat to a cold patient]), it is possible for the system to be over-gained, a situation in which an overly aggressive response destabilizes the system. Conversely, an undergained system may never exert control of the measured variable because the response is not effective enough. Ultimately, the stability of a simplified feedback loop system is a function of the inherent variability in the measured variable, response time, and gain.

Although, on the surface, this may seem to have only academic value, a thorough understanding of feedback loops is of great relevance to the anesthesiologist. For instance, an anesthesiologist who only administers 5- μ g doses of phenylephrine may never be able to exert meaningful control over BP (nor will the anesthesiologist who administers 500 μ g at a time). Similarly, BP stability is more easily achieved when near-instantaneous measures are used, as opposed to an oscillometric cuff that makes periodic measurements every 5 minutes. Also important is an appreciation for the lag time associated with these interventions; phenylephrine, for instance, takes 42 seconds to achieve peak effect.¹¹⁴ Appreciation for these basic engineering principles among anesthesiologists is not widespread, and they are not part of the American Board of Anesthesiology content outline. Nonetheless, basic instruction on the principles of feedback loops may be a high-yield endeavor for anesthesiology trainees.

Investigators at the University of California, Irvine, have taken the concept of feedback loops one step further and are attempting to “close the loop” using algorithms, which mimic the decision-making process used by individual anesthesiologists. The basic argument for this approach is that if some aspects of hemodynamic management can be easily standardized (e.g., giving IV fluid based on preestablished triggers), an **automated feedback system** that constantly monitors the measured variable (in this case, BP) will **outperform a human being by minimizing response time**, in addition to freeing up the anesthesiologist to focus on more cognitively challenging tasks. Preliminary work by this group has demonstrated feasibility and safety, but prospective comparisons between traditional, human-directed hemodynamic management and closed-loop hemodynamic management have yet to be undertaken.^{115,116}

CONCLUSIONS

Periodic, quantitative measurement of BP in humans, predating the era of evidence-based medicine by over a century, is a component of the American Society of Anesthesiologists standards for basic anesthetic monitoring and is a staple of anesthetic management worldwide. Thus, it is unlikely that the clinical value of BP monitoring will ever be assessed in prospective fashion. Furthermore, adherence to traditional BP parameters complicates the ability of investigators to determine whether particular BP ranges confer any clinical benefits. The **BP waveform** is a **complex amalgamation** of both **antegrade and retrograde (reflected) pressure waves** and is affected by vascular **compliance, distance** from the left ventricle, and the **3D structure** of the **vascular tree**. Accurate reproduction of this waveform requires an appropriately engineered monitoring system with a **frequency response of up to 40 Hz**. Oscillometry is the standard method of measuring BP semicontinuously in anesthetized patients and is the primary form of measurement in >80% of general anesthetics. Although these devices perform reasonably well when true MAP is approximately 75 mm Hg, a **major shortcoming of oscillometry** is its **poor performance** at the extremes. Another shortcoming of these devices is that they offer **no information about the shape of the BP waveform**. Two classes of devices have been developed to measure the BP waveform continuously, without requiring the placement of a catheter. **Tonometric devices “touch” the radial artery** by applying external pressure to the wrist. **Volume clamp devices combine a rapidly responsive finger cuff to a finger plethysmograph** in an effort to keep the radial artery in an “unstretched” state and the PPG flat; the **finger cuff pressure required to do so closely approximates the radial artery pressure**. The limits-of-agreement analyses of the latter 2 device classes using invasive measures as a reference standard are promising. Arterial catheterization remains the gold standard for accurate BP measurement, and understanding the measurement system is necessary for proper interpretation of the waveforms. In particular, the frequency response and damping coefficient of the fluid-filled tubing system can be used to measure the dynamic response of the system, and thus

its adequacy for the accurate reproduction of complex waveforms. Fortunately, both the **frequency response** and the **damping coefficient** can be **measured** at the bedside using the **flush test**. Control of BP requires measurement, an algorithm (usually human-controlled) and an intervention. This can be described as a **feedback loop**, of which **gain and response times** are important variables that affect the stability of the measured variable (in this case, BP). Several investigators have begun “closing” this loop, effectively automating the human algorithms used to manage BP, allowing the anesthesia providers to focus their intellectual energy on other, more complex decisions and tasks. ■

DISCLOSURES

Name: Karsten Bartels, MD.

Contribution: This author helped write and edit the manuscript.

Attestation: Karsten Bartels approved the final version of the manuscript.

Name: Stephen A. Esper, MD, MBA.

Contribution: This author helped write and edit the manuscript.

Attestation: Stephen A. Esper approved the final version of the manuscript.

Name: Robert H. Thiele, MD.

Contribution: This author helped write and edit the manuscript.

Attestation: Robert H. Thiele approved the final version of the manuscript.

This manuscript was handled by: Maxime Cannesson, MD, PhD.

REFERENCES

1. Oxman AD, Cook DJ, Guyatt GH. Users' guides to the medical literature. VI. How to use an overview. Evidence-Based Medicine Working Group. JAMA 1994;272:1367–71
2. Booth J. A short history of blood pressure measurement. Proc R Soc Med 1977;70:793–9
3. O'Rourke MF. The arterial pulse in health and disease. Am Heart J 1971;82:687–702
4. Parker KH. A brief history of arterial wave mechanics. Med Biol Eng Comput 2009;47:111–8
5. O'Brien E, Fitzgerald D. The history of blood pressure measurement. J Hum Hypertens 1994;8:73–84
6. Thiele RH, Nemergut EC, Lynch C III. The physiologic implications of isolated alpha-1 adrenergic stimulation. Anesth Analg 2011;113:284–96
7. Thiele RH, Nemergut EC, Lynch C III. The clinical implications of isolated alpha(1) adrenergic stimulation. Anesth Analg 2011;113:297–304
8. Reinhart K. [Monitoring of oxygen transport systems]. Anaesthesist 1988;37:1–9
9. Fog M. The relationship between the blood pressure and the tonic regulation of the pial arteries. J Neurol Psychiatry 1938;1:187–97
10. Gold JP, Charlson ME, Williams-Russo P, Szatrowski TP, Peterson JC, Pirraglia PA, Hartman GS, Yao FS, Hollenberg JP, Barbut D, Hayes JG, Thomas SJ, Purcell MH, Mattis S, Gorkin L, Post M, Krieger KH, Isom OW. Improvement of outcomes after coronary artery bypass. A randomized trial comparing intraoperative high versus low mean arterial pressure. J Thorac Cardiovasc Surg 1995;110:1302–11
11. Siepe M, Pfeiffer T, Gieringer A, Zemann S, Benk C, Schlensak C, Beyersdorf F. Increased systemic perfusion pressure during cardiopulmonary bypass is associated with less early postoperative cognitive dysfunction and delirium. Eur J Cardiothorac Surg 2011;40:200–7
12. Asfar P, Meziani F, Hamel JF, Grelon F, Megarbane B, Anguel N, Mira JP, Dequin PF, Gergaud S, Weiss N, Legay F, Le Tulzo Y, Conrad M, Robert R, Gonzalez F, Guitton C, Tamion F,

- Tonnellier JM, Guezennec P, Van Der Linden T, Vieillard-Baron A, Mariotte E, Pradel G, Lesieur O, Ricard JD, Hervé F, du Cheyron D, Guerin C, Mercat A, Teboul JL, Radermacher P; SEPSISPAM Investigators. High versus low blood-pressure target in patients with septic shock. *N Engl J Med* 2014;370:1583–93
13. Rivers E, Nguyen B, Havstad S, Ressler J, Muzzin A, Knoblich B, Peterson E, Tomlanovich M; Early Goal-Directed Therapy Collaborative Group. Early goal-directed therapy in the treatment of severe sepsis and septic shock. *N Engl J Med* 2001;345:1368–77
 14. Investigators A, Group ACT, Peake SL, Delaney A, Bailey M, Bellomo R, Cameron PA, Cooper DJ, Higgins AM, Holdgate A, Howe BD, Webb SA, Williams P. Goal-directed resuscitation for patients with early septic shock. *N Engl J Med*. 2014;371:1496–506
 15. Yealy DM, Kellum JA, Huang DT, Barnato AE, Weissfeld LA, Pike F, Terndrup T, Wang HE, Hou PC, LoVecchio F, Filbin MR, Shapiro NI, Angus DC; ProCESS Investigators. A randomized trial of protocol-based care for early septic shock. *N Engl J Med* 2014;370:1683–93
 16. Mouncey PR, Osborn TM, Power GS, Harrison DA, Sadique MZ, Grieve RD, Jahan R, Harvey SE, Bell D, Bion JF, Coats TJ, Singer M, Young JD, Rowan KM; ProMISe Trial Investigators. Trial of early, goal-directed resuscitation for septic shock. *N Engl J Med* 2015;372:1301–11
 17. Hamilton MA, Cecconi M, Rhodes A. A systematic review and meta-analysis on the use of preemptive hemodynamic intervention to improve postoperative outcomes in moderate and high-risk surgical patients. *Anesth Analg* 2011;112:1392–402
 18. Gurgel ST, do Nascimento P Jr. Maintaining tissue perfusion in high-risk surgical patients: a systematic review of randomized clinical trials. *Anesth Analg* 2011;112:1384–91
 19. Sandham JD, Hull RD, Brant RF, Knox L, Pineo GF, Doig CJ, Laporta DP, Viner S, Passerini L, Devitt H, Kirby A, Jacka M; Canadian Critical Care Clinical Trials Group. A randomized, controlled trial of the use of pulmonary-artery catheters in high-risk surgical patients. *N Engl J Med* 2003;348:5–14
 20. Bartels K, Thiele RH, Gan TJ. Rational fluid management in today's ICU practice. *Crit Care* 2013;17(suppl 1):S6
 21. Thiele RH, Isbell JM. Vasopressors and Inotropes (Chapter 11). In: Cannesson M, Pearse R, eds. *Perioperative Hemodynamic Monitoring and Goal Directed Therapy*. Cambridge: Cambridge University Press, 2014:85–94
 22. Sessler DI, Sigl JC, Kelley SD, Chamoun NG, Manberg PJ, Saager L, Kurz A, Greenwald S. Hospital stay and mortality are increased in patients having a “triple low” of low blood pressure, low bispectral index, and low minimum alveolar concentration of volatile anesthesia. *Anesthesiology* 2012;116:1195–203
 23. Walsh M, Devereaux PJ, Garg AX, Kurz A, Turan A, Rodseth RN, Cywinski J, Thabane L, Sessler DI. Relationship between intraoperative mean arterial pressure and clinical outcomes after noncardiac surgery: toward an empirical definition of hypotension. *Anesthesiology* 2013;119:507–15
 24. Warner HR, Swan HJ, Connolly DC, Tompkins RG, Wood EH. Quantitation of beat-to-beat changes in stroke volume from the aortic pulse contour in man. *J Appl Physiol* 1953;5:495–507
 25. Kouchoukos NT, Sheppard LC, McDonald DA. Estimation of stroke volume in the dog by a pulse contour method. *Circ Res* 1970;26:611–23
 26. Guier WH, Friesinger GC, Ross RS. Beat-by-beat stroke volume from aortic-pulse-pressure analysis. *IEEE Trans Biomed Eng* 1974;21:285–92
 27. Essler S, Schroeder MJ, Phaniraj V, Koenig SC, Latham RD, Ewert D. Fast estimation of arterial vascular parameters for transient and steady beats with application to hemodynamic state under variant gravitational conditions. *Ann Biomed Eng* 1999;27:486–97
 28. Laskey WK, Kussmaul WG. Arterial wave reflection in heart failure. *Circulation* 1987;75:711–22
 29. O'Rourke MF, Kelly RP. Wave reflection in the systemic circulation and its implications in ventricular function. *J Hypertens* 1993;11:327–37
 30. Cohn JN. Blood pressure measurement in shock. Mechanism of inaccuracy in auscultatory and palpatory methods. *JAMA* 1967;199:118–22
 31. Gould BA, Hornung RS, Kieso HA, Altman DG, Cashman PM, Raftery EB. Evaluation of the Remler M2000 blood pressure recorder. Comparison with intraarterial blood pressure recordings both at hospital and at home. *Hypertension* 1984;6:209–15
 32. Marey ÉJ. Recherches Sur L'état De La Circulation D'après Les Caractères Du Pouls Fournis Par Le Nouveau Sphygmographe. *J Physiol Homme Animaux* 1860;3:244
 33. Yelderman M, Ream AK. Indirect measurement of mean blood pressure in the anesthetized patient. *Anesthesiology* 1979;50:253–6
 34. Mauck GW, Smith CR, Geddes LA, Bourland JD. The meaning of the point of maximum oscillations in cuff pressure in the indirect measurement of blood pressure—part ii. *J Biomech Eng* 1980;102:28–33
 35. Posey JA, Geddes LA, Williams H, Moore AG. The meaning of the point of maximum oscillations in cuff pressure in the indirect measurement of blood pressure. 1. *Cardiovasc Res Cent Bull* 1969;8:15–25
 36. Peñáz J. Photoelectric Measurement of Blood Pressure, Volume and Flow in the Finger. *Digest 10th Int Conf Med Biol Eng*, 104, Dresden, Germany, 1973
 37. Pickering TG, Hall JE, Falkner BE, Graves J, Hill MN, Jones DW, Kurtz T, Sheps SG, Roccella EJ; Subcommittee of Professional and Public Education of the American Heart Association Council on High Blood Pressure Research. Recommendations for blood pressure measurement in humans and experimental animals: Part 1: blood pressure measurement in humans: a statement for professionals from the Subcommittee of Professional and Public Education of the American Heart Association Council on High Blood Pressure Research. *Hypertension* 2005;45:142–61
 38. Park MK, Lee DH, Johnson GA. Oscillometric blood pressures in the arm, thigh, and calf in healthy children and those with aortic coarctation. *Pediatrics* 1993;91:761–5
 39. Stebor AD. Basic principles of noninvasive blood pressure measurement in infants. *Adv Neonatal Care* 2005;5:252–61
 40. Wax DB, Lin HM, Leibowitz AB. Invasive and concomitant non-invasive intraoperative blood pressure monitoring: observed differences in measurements and associated therapeutic interventions. *Anesthesiology* 2011;115:973–8
 41. Kim SH, Lilot M, Sidhu KS, Rinehart J, Yu Z, Canales C, Cannesson M. Accuracy and precision of continuous non-invasive arterial pressure monitoring compared with invasive arterial pressure: a systematic review and meta-analysis. *Anesthesiology* 2014;120:1080–97
 42. Heard SO, Lisbon A, Toth I, Ramasubramanian R. An evaluation of a new continuous blood pressure monitoring system in critically ill patients. *J Clin Anesth* 2000;12:509–18
 43. Borow KM, Newburger JW. Noninvasive estimation of central aortic pressure using the oscillometric method for analyzing systemic artery pulsatile blood flow: comparative study of indirect systolic, diastolic, and mean brachial artery pressure with simultaneous direct ascending aortic pressure measurements. *Am Heart J* 1982;103:879–86
 44. Loubser PG. Comparison of intra-arterial and automated oscillometric blood pressure measurement methods in postoperative hypertensive patients. *Med Instrum* 1986;20:255–9
 45. Gravlee GP, Brockschmidt JK. Accuracy of four indirect methods of blood pressure measurement, with hemodynamic correlations. *J Clin Monit* 1990;6:284–98
 46. Gorbach MS, Quill TJ, Lavine ML. The relative accuracies of two automated noninvasive arterial pressure measurement devices. *J Clin Monit* 1991;7:13–22
 47. Manolio TA, Fishel SC, Beattie C, Torres J, Christopherson R, Merritt WT, Whelton PK. Evaluation of the Dinamap continuous blood pressure monitor. *Am J Hypertens* 1988;1:161S–7S
 48. O'Rourke MF, Pauca A, Jiang XJ. Pulse wave analysis. *Br J Clin Pharmacol* 2001;51:507–22
 49. Kemmotsu O, Ueda M, Otsuka H, Yamamura T, Winter DC, Eckerle JS. Arterial tonometry for noninvasive, continuous

- blood pressure monitoring during anesthesia. *Anesthesiology* 1991;75:333–40
50. Janelle GM, Gravenstein N. An accuracy evaluation of the T-Line Tensometer (continuous noninvasive blood pressure management device) versus conventional invasive radial artery monitoring in surgical patients. *Anesth Analg* 2006;102:484–90
51. Szmuk P, Pivalizza E, Warters RD, Ezri T, Gebhard R. An evaluation of the T-Line Tensometer continuous noninvasive blood pressure device during induced hypotension. *Anaesthesia* 2008;63:307–12
52. Dueck R, Goedje O, Clopton P. Noninvasive continuous beat-to-beat radial artery pressure via TL-200 applanation tonometry. *J Clin Monit Comput* 2012;26:75–83
53. Colquhoun DA, Forkin KT, Dunn LK, Bogdonoff DL, Durieux ME, Thiele RH. Non-invasive, minute-to-minute estimates of systemic arterial pressure and pulse pressure variation using radial artery tonometry. *J Med Eng Technol* 2013;37:197–202
54. Hamilton W, Remington J, Dow P. The determination of the propagation velocity of the arterial pulse wave. *Am J Physiol* 1945;144:521–35
55. Greene ES, Gerson JI. Arterial pulse wave velocity: a limited index of systemic vascular resistance during normotensive anesthesia in dogs. *J Clin Monit* 1985;1:219–26
56. Talke P, Nichols RJ Jr, Traber DL. Does measurement of systolic blood pressure with a pulse oximeter correlate with conventional methods? *J Clin Monit* 1990;6:5–9
57. Young CC, Mark JB, White W, DeBree A, Vender JS, Fleming A. Clinical evaluation of continuous noninvasive blood pressure monitoring: accuracy and tracking capabilities. *J Clin Monit* 1995;11:245–52
58. Hirano H, Horiuchi T, Kutluk A, Kurita Y, Ukawa T, Nakamura R, Saeki N, Higashi Y, Kawamoto M, Yoshizumi M, Tsuji T. A log-linearized arterial viscoelastic model for evaluation of the carotid artery. *Conf Proc IEEE Eng Med Biol Soc* 2013;2013:2591–4
59. Kim SH, Song JG, Park JH, Kim JW, Park YS, Hwang GS. Beat-to-beat tracking of systolic blood pressure using noninvasive pulse transit time during anesthesia induction in hypertensive patients. *Anesth Analg* 2013;116:94–100
60. Sharwood-Smith G, Bruce J, Drummond G. Assessment of pulse transit time to indicate cardiovascular changes during obstetric spinal anaesthesia. *Br J Anaesth* 2006;96:100–5
61. Jilek J, Stork M. Oscillometric pressure pulse waveforms: their current and prospective applications in biomedical instrumentation. *Proceedings of the 13th WSEAS International Conference on SYSTEMS* 2009:133–8
62. Bogert LW, van Lieshout JJ. Non-invasive pulsatile arterial pressure and stroke volume changes from the human finger. *Exp Physiol* 2005;90:437–46
63. Imholz BP, Wieling W, van Montfrans GA, Wesseling KH. Fifteen years experience with finger arterial pressure monitoring: assessment of the technology. *Cardiovasc Res* 1998;38:605–16
64. Saugel B, Dueck R, Wagner JY. Measurement of blood pressure. *Best Pract Res Clin Anaesthesiol* 2014;28:309–22
65. Bos WJ, van Goudoever J, van Montfrans GA, van den Meiracker AH, Wesseling KH. Reconstruction of brachial artery pressure from noninvasive finger pressure measurements. *Circulation* 1996;94:1870–5
66. Wesseling KH. Finger arterial pressure measurement with Finapres. *Z Kardiol* 1996;85(suppl 3):38–44
67. Fortin J, Grillenberger R. Apparatus and Method for Enhancing and Analyzing Signals from a Continuous Non-Invasive Blood Pressure Device. Graz, Austria: Cnsystems Medizintechnik Ag, 2011
68. Bartels K, Thiele RH. Advances in photoplethysmography: beyond arterial oxygen saturation. *Can J Anaesth* 2015;62:1313–28
69. Awad AA, Ghashy MA, Ouda W, Stout RG, Silverman DG, Shelley KH. Different responses of ear and finger pulse oximeter wave form to cold pressor test. *Anesth Analg* 2001;92:1483–6
70. Dorlas JC, Nijboer JA, Butijn WT, van der Hoeven GM, Settels JJ, Wesseling KH. Effects of peripheral vasoconstriction on the blood pressure in the finger, measured continuously by a new noninvasive method (the Finapres). *Anesthesiology* 1985;62:342–5
71. Kurki T, Smith NT, Head N, Dec-Silver H, Quinn A. Noninvasive continuous blood pressure measurement from the finger: optimal measurement conditions and factors affecting reliability. *J Clin Monit* 1987;3:6–13
72. Stover JF, Stocker R, Lenherr R, Neff TA, Cottini SR, Zoller B, Béchir M. Noninvasive cardiac output and blood pressure monitoring cannot replace an invasive monitoring system in critically ill patients. *BMC Anesthesiol* 2009;9:6
73. Biais M, Vidil L, Rouillet S, Masson F, Quinart A, Revel P, Szark F. Continuous non-invasive arterial pressure measurement: evaluation of CNAP device during vascular surgery. *Ann Fr Anesth Reanim* 2010;29:530–5
74. Hofhuizen CM, Lemson J, Hemelaar AE, Settels JJ, Schraa O, Singh SK, van der Hoeven JG, Scheffer GJ. Continuous non-invasive finger arterial pressure monitoring reflects intra-arterial pressure changes in children undergoing cardiac surgery. *Br J Anaesth* 2010;105:493–500
75. Jeleazcov C, Krajcinovic L, Münster T, Birkholz T, Fried R, Schüttler J, Fechner J. Precision and accuracy of a new device (CNAP™) for continuous non-invasive arterial pressure monitoring: assessment during general anaesthesia. *Br J Anaesth* 2010;105:264–72
76. Martina JR, Westerhof BE, Van Goudoever J, De Jonge N, Van Lieshout JJ, Lahpor JR, De Mol BA. Noninvasive blood pressure measurement by the Nexfin monitor during reduced arterial pulsatility: a feasibility study. *ASAIO J* 2010;56:221–7
77. Schramm C, Baat L, Plaschke K. Continuous noninvasive arterial pressure: assessment in older and high-risk patients under analgesic sedation. *Blood Press Monit* 2011;16:270–6
78. Fischer MO, Avram R, Cârjalu I, Massetti M, Gérard JL, Hanouz JL, Fellahi JL. Non-invasive continuous arterial pressure and cardiac index monitoring with Nexfin after cardiac surgery. *Br J Anaesth* 2012;109:514–21
79. Garnier RP, van der Spoel AG, Sibarani-Ponsen R, Markhorst DG, Boer C. Level of agreement between Nexfin non-invasive arterial pressure with invasive arterial pressure measurements in children. *Br J Anaesth* 2012;109:609–15
80. Hahn R, Rinösl H, Neuner M, Kettner SC. Clinical validation of a continuous non-invasive haemodynamic monitor (CNAP™; 500) during general anaesthesia. *Br J Anaesth* 2012;108:581–5
81. Ilies C, Bauer M, Berg P, Rosenberg J, Hedderich J, Bein B, Hinz J, Hanss R. Investigation of the agreement of a continuous non-invasive arterial pressure device in comparison with invasive radial artery measurement. *Br J Anaesth* 2012;108:202–10
82. Jagadeesh AM, Singh NG, Mahankali S. A comparison of a continuous noninvasive arterial pressure (CNAP™) monitor with an invasive arterial blood pressure monitor in the cardiac surgical ICU. *Ann Card Anaesth* 2012;15:180–4
83. Martina JR, Westerhof BE, van Goudoever J, de Beaumont EM, Truijen J, Kim YS, Immink RV, Jöbsis DA, Hollmann MW, Lahpor JR, de Mol BA, van Lieshout JJ. Noninvasive continuous arterial blood pressure monitoring with Nexfin®. *Anesthesiology* 2012;116:1092–103
84. Broch O, Bein B, Gruenewald M, Carstens A, Illies C, Schöneich F, Steinfath M, Renner J. A comparison of continuous non-invasive arterial pressure with invasive radial and femoral pressure in patients undergoing cardiac surgery. *Minerva Anesthesiol* 2013;79:248–56
85. Dewhurst E, Corridore M, Klamar J, Beebe A, Rice J, Barry N, Tobias JD. Accuracy of the CNAP monitor, a noninvasive continuous blood pressure device, in providing beat-to-beat blood pressure readings in the prone position. *J Clin Anesth* 2013;25:309–13
86. Gayat E, Mongardon N, Tuil O, Sievert K, Chazot T, Liu N, Fischler M. CNAP® does not reliably detect minimal or maximal arterial blood pressures during induction of anaesthesia and tracheal intubation. *Acta Anaesthesiol Scand* 2013;57:468–73
87. Hohn A, Defosse JM, Becker S, Steffen C, Wappler F, Sakka SG. Non-invasive continuous arterial pressure monitoring with Nexfin does not sufficiently replace invasive measurements in critically ill patients. *Br J Anaesth* 2013;111:178–84
88. Kako H, Corridore M, Rice J, Tobias JD. Accuracy of the CNAP™ monitor, a noninvasive continuous blood pressure

- device, in providing beat-to-beat blood pressure readings in pediatric patients weighing 20-40 kilograms. *Paediatr Anaesth* 2013;23:989-93
89. Schramm C, Huber A, Plaschke K. The accuracy and responsiveness of continuous noninvasive arterial pressure during rapid ventricular pacing for transcatheter aortic valve replacement. *Anesth Analg* 2013;117:76-82
 90. Ameloot K, Van De Vijver K, Van Regenmortel N, De Laet I, Schoonheydt K, Dits H, Broch O, Bein B, Malbrain ML. Validation study of Nexfin® continuous non-invasive blood pressure monitoring in critically ill adult patients. *Minerva Anesthesiol* 2014;80:1294-301
 91. Ilies C, Grudev G, Hedderich J, Renner J, Steinfath M, Bein B, Haake N, Hanss R. Comparison of a continuous noninvasive arterial pressure device with invasive measurements in cardiovascular postsurgical intensive care patients: a prospective observational study. *Eur J Anaesthesiol* 2015;32:20-8
 92. Kumar GA, Jagadeesh AM, Singh NG, Prasad SR. Evaluation of continuous non-invasive arterial pressure monitoring during induction of general anaesthesia in patients undergoing cardiac surgery. *Indian J Anaesth* 2015;59:21-5
 93. Tobias JD, McKee C, Herz D, Teich S, Sohner P, Rice J, Barry N, Michalsky M. Accuracy of the CNAP™ monitor, a noninvasive continuous blood pressure device, in providing beat-to-beat blood pressure measurements during bariatric surgery in severely obese adolescents and young adults. *J Anesth* 2014;28:861-5
 94. Vos JJ, Poterman M, Mooyaart EA, Weening M, Struys MM, Scheeren TW, Kalmar AF. Comparison of continuous non-invasive finger arterial pressure monitoring with conventional intermittent automated arm arterial pressure measurement in patients under general anaesthesia. *Br J Anaesth* 2014;113:67-74
 95. Wagner JY, Negulescu I, Schöffthaler M, Hapfelmeier A, Meidert AS, Huber W, Schmid RM, Saugel B. Continuous noninvasive arterial pressure measurement using the volume clamp method: an evaluation of the CNAP device in intensive care unit patients. *J Clin Monit Comput* 2015;29:807-13
 96. Weiss E, Gayat E, Dumans-Nizard V, Le Guen M, Fischler M. Use of the Nexfin™ device to detect acute arterial pressure variations during anaesthesia induction. *Br J Anaesth* 2014;113:52-60
 97. Maguire S, Rinehart J, Vakharia S, Cannesson M. Technical communication: respiratory variation in pulse pressure and plethysmographic waveforms: intraoperative applicability in a North American academic center. *Anesth Analg* 2011;112:94-6
 98. Esper SA, Pinsky MR. Arterial waveform analysis. *Best Pract Res Clin Anaesthesiol* 2014;28:363-80
 99. Thiele RH, Durieux ME. Arterial waveform analysis for the anesthesiologist: past, present, and future concepts. *Anesth Analg* 2011;113:766-76
 100. Thiele RH. Cardiac bulldozers, backhoes, and blood pressure. *Anesth Analg* 2015;121:1417-9
 101. Handlogten KS, Wilson GA, Clifford L, Nuttall GA, Kor DJ. Brachial artery catheterization: an assessment of use patterns and associated complications. *Anesth Analg* 2014;118:288-95
 102. Maki DG, Kluger DM, Crnich CJ. The risk of bloodstream infection in adults with different intravascular devices: a systematic review of 200 published prospective studies. *Mayo Clin Proc* 2006;81:1159-71
 103. Scheer B, Perel A, Pfeiffer UJ. Clinical review: complications and risk factors of peripheral arterial catheters used for haemodynamic monitoring in anaesthesia and intensive care medicine. *Crit Care* 2002;6:199-204
 104. Levin PD, Sheinin O, Gozal Y. Use of ultrasound guidance in the insertion of radial artery catheters. *Crit Care Med* 2003;31:481-4
 105. Tang L, Wang F, Li Y, Zhao L, Xi H, Guo Z, Li X, Gao C, Wang J, Zhou L. Ultrasound guidance for radial artery catheterization: an updated meta-analysis of randomized controlled trials. *PLoS One* 2014;9:e111527
 106. Wilkins RG. Radial artery cannulation and ischaemic damage: a review. *Anaesthesia* 1985;40:896-9
 107. Slogoff S, Keats AS, Arlund C. On the safety of radial artery cannulation. *Anesthesiology* 1983;59:42-7
 108. Barbeau GR, Arsenault F, Dugas L, Simard S, Larivière MM. Evaluation of the ulnopalmar arterial arches with pulse oximetry and plethysmography: comparison with the Allen's test in 1010 patients. *Am Heart J* 2004;147:489-93
 109. Abu-Omar Y, Mussa S, Anastasiadis K, Steel S, Hands L, Taggart DP. Duplex ultrasonography predicts safety of radial artery harvest in the presence of an abnormal Allen test. *Ann Thorac Surg* 2004;77:116-9
 110. Gardner RM. Direct blood pressure measurement—dynamic response requirements. *Anesthesiology* 1981;54:227-36
 111. Thiele RH, Colquhoun DA, Tucker-Schwartz JM, Gillies GT, Durieux ME. Radial-femoral concordance in time and frequency domain-based estimates of systemic arterial respiratory variation. *J Clin Monit Comput* 2012;26:393-400
 112. Timsit JF, Schwebel C, Bouadma L, Geffroy A, Garrouste-Orgeas M, Pease S, Herauld MC, Haouache H, Calvino-Gunther S, Gustin B, Armand-Lefevre L, Leflon V, Chaplain C, Benali A, Francois A, Adrie C, Zahar JR, Thuong M, Arrault X, Croize J, Lucet JC; Dressing Study Group. Chlorhexidine-impregnated sponges and less frequent dressing changes for prevention of catheter-related infections in critically ill adults: a randomized controlled trial. *JAMA* 2009;301:1231-41
 113. Marlin TE. *Process Control: Designing Processes and Control Systems for Dynamic Performance*. 2nd ed. New York, NY: McGraw Hill, 2000
 114. Schwinn DA, Reves JG. Time course and hemodynamic effects of alpha-1-adrenergic bolus administration in anesthetized patients with myocardial disease. *Anesth Analg* 1989;68:571-8
 115. Rinehart J, Lilot M, Lee C, Joosten A, Huynh T, Canales C, Imagawa D, Demirjian A, Cannesson M. Closed-loop assisted versus manual goal-directed fluid therapy during high-risk abdominal surgery: a case-control study with propensity matching. *Crit Care* 2015;19:94
 116. Rinehart J, Le Manach Y, Douiri H, Lee C, Lilot M, Le K, Canales C, Cannesson M. First closed-loop goal directed fluid therapy during surgery: a pilot study. *Ann Fr Anesth Reanim* 2014;33:e35-41
 117. Briet M, Boutouyrie P, Laurent S, London GM. Arterial stiffness and pulse pressure in CKD and ESRD. *Kidney International* 2012;82:388-400
 118. Geddes LA. *Cardiovascular Devices And Their Applications*. Chapter 3, "Blood Pressure: Noninvasive Measurement," p 72